

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032124\
 Data File : VW034778.D
 Acq On : 21 Mar 2024 10:49
 Operator : SY/MD
 Sample : P1778-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 22 02:00:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 22 01:59:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	337310	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	317627	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	150483	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	158988	54.371	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 108.740%		
7) Chloroethane-d5	1.532	69	125956	55.290	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 110.580%		
11) 1,1-Dichloroethene-d2	2.060	65	58599	44.312	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 88.620%		
21) 2-Butanone-d5	3.793	46	186233	95.237	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 95.240%		
24) Chloroform-d	4.249	84	264281	51.117	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.240%		
26) 1,2-Dichloroethane-d4	4.941	65	179085	53.006	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.020%		
32) Benzene-d6	4.957	84	522145	54.469	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.940%		
36) 1,2-Dichloropropane-d6	5.986	67	157356	53.788	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 107.580%		
41) Toluene-d8	7.243	98	462429	53.306	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.620%		
43) trans-1,3-Dichloroprop...	7.551	79	74761	50.889	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.780%		
47) 2-Hexanone-d5	8.024	63	110606	75.713	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 75.710%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	199963	51.940	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 103.880%		
66) 1,2-Dichlorobenzene-d4	11.558	152	167730	54.530	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 109.060%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032124\
Data File : VV034778.D
Acq On : 21 Mar 2024 10:49
Operator : SY/MD
Sample : P1778-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 22 02:00:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 22 01:59:49 2024
Response via : Initial Calibration

