

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW032124\  
 Data File : VW034780.D  
 Acq On : 21 Mar 2024 11:38  
 Operator : SY/MD  
 Sample : P1743-01 10X  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 E0732

Quant Time: Mar 22 02:01:11 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM022924WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Fri Mar 22 01:59:49 2024  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon           | Response | Conc       | Units | Dev(Min) |
|-------------------------------|---------|----------------|----------|------------|-------|----------|
| -----                         |         |                |          |            |       |          |
| Internal Standards            |         |                |          |            |       |          |
| 1) 1,4-Difluorobenzene        | 5.532   | 114            | 347814   | 50.000     | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.783   | 117            | 331457   | 50.000     | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.185  | 152            | 157751   | 50.000     | ug/L  | 0.00     |
| System Monitoring Compounds   |         |                |          |            |       |          |
| 4) Vinyl Chloride-d3          | 1.282   | 65             | 156412   | 51.875     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 135 | Recovery | = 103.740% |       |          |
| 7) Chloroethane-d5            | 1.532   | 69             | 122148   | 51.999     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 130 | Recovery | = 104.000% |       |          |
| 11) 1,1-Dichloroethene-d2     | 2.060   | 65             | 49057    | 35.976     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 125 | Recovery | = 71.960%  |       |          |
| 21) 2-Butanone-d5             | 3.789   | 46             | 161434   | 80.062     | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range 40 - 130 | Recovery | = 80.060%  |       |          |
| 24) Chloroform-d              | 4.246   | 84             | 255414   | 47.910     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | = 95.820%  |       |          |
| 26) 1,2-Dichloroethane-d4     | 4.937   | 65             | 173574   | 49.823     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | = 99.640%  |       |          |
| 32) Benzene-d6                | 4.957   | 84             | 506265   | 50.609     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | = 101.220% |       |          |
| 36) 1,2-Dichloropropane-d6    | 5.986   | 67             | 156949   | 51.410     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 120 | Recovery | = 102.820% |       |          |
| 41) Toluene-d8                | 7.239   | 98             | 457108   | 50.494     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 80 - 120 | Recovery | = 100.980% |       |          |
| 43) trans-1,3-Dichloroprop... | 7.551   | 79             | 75273    | 49.100     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 125 | Recovery | = 98.200%  |       |          |
| 47) 2-Hexanone-d5             | 8.024   | 63             | 99778    | 65.451     | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range 45 - 130 | Recovery | = 65.450%  |       |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.149  | 84             | 190020   | 47.298     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 65 - 120 | Recovery | = 94.600%  |       |          |
| 66) 1,2-Dichlorobenzene-d4    | 11.558  | 152            | 169416   | 52.541     | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range 80 - 120 | Recovery | = 105.080% |       |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV032124\  
Data File : VV034780.D  
Acq On : 21 Mar 2024 11:38  
Operator : SY/MD  
Sample : P1743-01 10X  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
E0732

Quant Time: Mar 22 02:01:11 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM022924WMA.M  
Quant Title : VOC Analysis  
QLast Update : Fri Mar 22 01:59:49 2024  
Response via : Initial Calibration

