

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032124\
 Data File : VV034783.D
 Acq On : 21 Mar 2024 12:50
 Operator : SY/MD
 Sample : P1755-15DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCKX4DL

Quant Time: Mar 22 02:01:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 22 01:59:49 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	332820	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	316502	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	148391	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	151981	52.676	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 105.360%		
7) Chloroethane-d5	1.532	69	121119	53.884	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 107.760%		
11) 1,1-Dichloroethene-d2	2.060	65	52118	39.942	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 79.880%		
21) 2-Butanone-d5	3.793	46	171631	88.954	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 88.950%		
24) Chloroform-d	4.246	84	254233	49.836	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.680%		
26) 1,2-Dichloroethane-d4	4.941	65	169408	50.818	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.640%		
32) Benzene-d6	4.957	84	509760	53.366	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.740%		
36) 1,2-Dichloropropane-d6	5.986	67	151353	51.920	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 103.840%		
41) Toluene-d8	7.239	98	447273	51.742	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 103.480%		
43) trans-1,3-Dichloroprop...	7.551	79	74033	50.573	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.140%		
47) 2-Hexanone-d5	8.024	63	111447	76.560	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 76.560%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	193884	50.540	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 101.080%		
66) 1,2-Dichlorobenzene-d4	11.558	152	160011	52.754	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.500%		
Target Compounds						
13) Acetone	2.121	43	197230	85.031	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032124\
Data File : VV034783.D
Acq On : 21 Mar 2024 12:50
Operator : SY/MD
Sample : P1755-15DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCKX4DL

Quant Time: Mar 22 02:01:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM022924WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 22 01:59:49 2024
Response via : Initial Calibration

