

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032219\
 Data File : VV009824.D
 Acq On : 22 Mar 2019 13:53
 Operator : SY/MD
 Sample : K2049-02MS
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 P002-TW001-01MS

Quant Time: Mar 23 04:23:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 23 03:53:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	111593	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	109151	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	46706	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54971	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	1.59	69	66610	5.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.60%
11) 1,1-Dichloroethene-d2	2.14	63	144279	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.20%
20) 2-Butanone-d5	3.97	46	74755	47.89	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.78%
24) Chloroform-d	4.41	84	61678	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
26) 1,2-Dichloroethane-d4	5.09	65	32954	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.10	84	128519	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	6.12	67	38958	4.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.20%
41) Toluene-d8	7.36	98	114674	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.80%
43) trans-1,3-Dichloropropene-	7.66	79	11719	3.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.40%
46) 2-Hexanone-d5	8.14	63	57580	42.92	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.84%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	26811	4.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	40685	4.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.00%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	65698	4.415 ug/L	92
33) Benzene	5.15	78	138171	4.507 ug/L	100
34) Trichloroethene	5.96	95	33964	4.380 ug/L	94
42) Toluene	7.43	91	147128	4.241 ug/L	97
51) Chlorobenzene	8.93	112	97637	4.396 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032219\
Data File : VV009824.D
Acq On : 22 Mar 2019 13:53
Operator : SY/MD
Sample : K2049-02MS
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
P002-TW001-01MS

Quant Time: Mar 23 04:23:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 23 03:53:22 2019
Response via : Initial Calibration

