

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032219\
 Data File : VV009839.D
 Acq On : 22 Mar 2019 20:16
 Operator : SY/MD
 Sample : K2052-14
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 P022-TW001-01

Quant Time: Mar 23 05:43:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 23 03:53:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	87608	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	89068	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	36323	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48831	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.59	69	50207	5.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.00%
11) 1,1-Dichloroethene-d2	2.13	63	94967	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.96	46	64220	52.41	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.82%
24) Chloroform-d	4.41	84	55216	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.08	65	27824	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.40%
32) Benzene-d6	5.10	84	112356	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
36) 1,2-Dichloropropane-d6	6.12	67	35878	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.36	98	95324	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.67	79	9503	3.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.80%
46) 2-Hexanone-d5	8.14	63	51522	47.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	24435	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	35610	5.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.00%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	3389	0.536 ug/L	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032219\
Data File : VV009839.D
Acq On : 22 Mar 2019 20:16
Operator : SY/MD
Sample : K2052-14
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
P022-TW001-01

Quant Time: Mar 23 05:43:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 23 03:53:22 2019
Response via : Initial Calibration

