

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032219\
 Data File : VV009827.D
 Acq On : 22 Mar 2019 15:09
 Operator : SY/MD
 Sample : K2052-16
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TB-190320-01

Quant Time: Mar 23 04:29:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 23 03:53:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	89421	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	90530	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	38347	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	52860	4.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.80%
7) Chloroethane-d5	1.59	69	50830	5.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.20%
11) 1,1-Dichloroethene-d2	2.13	63	102674	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.20%
20) 2-Butanone-d5	3.98	46	66663	53.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.60%
24) Chloroform-d	4.41	84	57007	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.09	65	29217	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.40%
32) Benzene-d6	5.10	84	119369	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.12	67	36430	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.36	98	107166	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloropropene-	7.67	79	10654	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.14	63	58488	52.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	25625	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	38089	5.47	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032219\
Data File : VV009827.D
Acq On : 22 Mar 2019 15:09
Operator : SY/MD
Sample : K2052-16
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TB-190320-01

Quant Time: Mar 23 04:29:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR032119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 23 03:53:22 2019
Response via : Initial Calibration

