

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032223\
 Data File : VW030296.D
 Acq On : 22 Mar 2023 16:43
 Operator : SY/MD
 Sample : 02003-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 23 01:50:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 23 01:48:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	114649	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	111980	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	58612	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	33707	4.544	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.800%
7) Chloroethane-d5	1.535	69	32595	4.681	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.600%
11) 1,1-Dichloroethene-d2	2.063	63	56799	3.724	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	3.825	46	61983	45.318	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.640%
24) Chloroform-d	4.262	84	70478	4.651	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	4.957	65	37014	5.983	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.600%
32) Benzene-d6	4.969	84	128116	4.252	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
36) 1,2-Dichloropropane-d6	5.998	67	35718	3.940	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.800%
41) Toluene-d8	7.252	98	117882	4.127	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	7.564	79	13625	4.393	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.800%
46) 2-Hexanone-d5	8.037	63	53962	42.057	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.120%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	25477	3.951	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	11.570	152	46927	4.732	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032223\
Data File : VV030296.D
Acq On : 22 Mar 2023 16:43
Operator : SY/MD
Sample : 02003-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 23 01:50:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 23 01:48:30 2023
Response via : Initial Calibration

