


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032223\
Data File : VV030289.D
Acq On : 22 Mar 2023 11:10
Operator : SY/MD
Sample : BFB237
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB237

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR022323WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Thu Mar 23 01:48:30 2023



AutoFind: Scans 2792, 2793, 2794; Background Corrected with Scan 2780

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	5032	PASS
75	95	30	80	52.4	15333	PASS
95	95	100	100	100.0	29237	PASS
96	95	5	9	6.9	2022	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	103.3	30189	PASS
175	174	5	9	7.2	2164	PASS
176	174	95	105	96.3	29085	PASS
177	176	5	10	6.5	1883	PASS

m/z	Abundance
39.90	156.0
43.90	126.0

Instrument :
MSVOA_V
ClientSampleId :
BFB237

m/z	Abundance
36.10	317.0
37.10	1406.0
38.00	1425.0
39.00	625.0
39.90	123.0
43.20	107.0
44.00	238.0
45.00	297.0
47.10	243.0
48.00	282.0
49.00	1151.0
50.00	5114.0
51.10	1518.0
52.00	105.0
52.70	104.0
55.00	132.0
56.10	436.0
57.10	885.0
60.00	390.0
61.00	1621.0
62.00	1265.0
63.10	1338.0
63.90	282.0
65.10	260.0
68.10	2885.0
69.10	3149.0
70.00	295.0
71.90	180.0
73.00	1627.0
74.10	5226.0
75.10	15603.0
76.00	1446.0
77.00	204.0
78.10	131.0
79.00	1226.0
79.90	335.0
80.90	1328.0
82.00	267.0
82.90	103.0
87.00	1291.0
87.90	910.0
91.00	145.0
92.00	827.0
93.00	1415.0
94.10	3718.0
95.10	29704.0
96.00	2119.0
103.90	173.0
105.80	236.0
114.70	100.0
116.00	151.0
116.90	291.0
117.90	149.0
118.90	209.0
128.00	160.0
130.00	142.0
131.00	105.0
134.90	111.0
140.90	428.0
142.90	453.0
144.80	126.0
148.00	123.0
154.90	107.0
170.90	101.0
171.80	177.0
174.00	29840.0
175.10	2144.0
176.00	28864.0
177.00	1903.0

Instrument :
MSVOA_V
ClientSampleId :
BFB237

m/z	Abundance
36.10	304.0
37.00	1413.0
38.00	1474.0
39.00	641.0
40.00	136.0
44.00	263.0
45.00	307.0
47.10	243.0
47.90	285.0
49.00	1152.0
50.10	5184.0
51.00	1594.0
52.80	130.0
55.00	163.0
56.10	461.0
57.10	888.0
57.80	125.0
60.00	394.0
61.00	1595.0
62.00	1378.0
63.10	1490.0
63.90	255.0
65.10	341.0
68.10	2963.0
69.10	3023.0
70.00	279.0
71.90	195.0
73.00	1621.0
74.10	5331.0
75.10	15882.0
76.00	1383.0
76.90	189.0
78.10	152.0
79.00	1228.0
79.90	366.0
80.90	1273.0
82.00	288.0
87.00	1284.0
87.90	899.0
91.00	145.0
92.00	832.0
93.10	1330.0
94.10	3650.0
95.10	30232.0
96.00	2113.0
103.80	205.0
105.80	268.0
116.00	155.0
116.90	340.0
117.90	197.0
118.80	259.0
128.00	151.0
130.00	108.0
131.10	101.0
135.00	111.0
140.90	436.0
142.90	454.0
144.90	127.0
154.90	132.0
170.90	125.0
171.90	247.0
174.00	31136.0
175.10	2230.0
176.00	30008.0
176.90	1925.0

Instrument :
MSVOA_V
ClientSampleId :
BFB237

m/z	Abundance
36.20	296.0
37.00	1269.0
38.00	1313.0
39.00	582.0
40.00	147.0
44.00	296.0
45.00	295.0
47.00	240.0
47.90	266.0
49.00	1096.0
50.10	4797.0
51.00	1491.0
52.00	114.0
52.90	112.0
55.00	135.0
56.00	460.0
57.00	808.0
57.80	122.0
60.00	329.0
61.00	1441.0
62.00	1423.0
63.10	1395.0
63.90	185.0
65.10	357.0
68.10	2725.0
69.00	2785.0
70.10	297.0
72.00	185.0
73.00	1382.0
74.00	4836.0
75.10	14515.0
76.00	1137.0
76.90	170.0
78.10	151.0
79.00	1076.0
79.90	364.0
80.90	1111.0
82.00	268.0
87.00	1108.0
87.90	857.0
91.00	123.0
92.00	832.0
93.10	1096.0
94.00	3335.0
95.10	27776.0
96.10	1834.0
103.80	182.0
105.80	247.0
116.00	137.0
116.90	327.0
118.00	198.0
118.80	284.0
128.10	125.0
135.00	108.0
141.00	400.0
142.80	437.0
145.00	168.0
154.90	116.0
170.90	108.0
172.00	283.0
174.00	29592.0
175.00	2117.0
176.00	28384.0
176.90	1821.0

Instrument :
MSVOA_V
ClientSampleId :
BFB237