

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032322\
 Data File : VW025135.D
 Acq On : 23 Mar 2022 16:20
 Operator : SY/MD
 Sample : N2083-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW5Z3

Quant Time: Mar 24 01:17:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 24 01:12:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	129712	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	122992	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55169	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	41103	4.014	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	80.200%	
7) Chloroethane-d5	1.571	69	36909	4.922	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	98.400%	
11) 1,1-Dichloroethene-d2	2.111	63	53571	3.237	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.800%	
20) 2-Butanone-d5	3.896	46	88736	71.700	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	143.400%#	
24) Chloroform-d	4.349	84	85360	5.026	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
26) 1,2-Dichloroethane-d4	5.037	65	37860	5.332	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.600%	
32) Benzene-d6	5.050	84	169849	4.761	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.200%	
36) 1,2-Dichloropropane-d6	6.069	67	52593	5.126	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.600%	
41) Toluene-d8	7.317	98	142338	4.377	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.600%	
43) trans-1,3-Dichloroprop...	7.629	79	14506	4.056	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.200%	
46) 2-Hexanone-d5	8.088	63	69137	56.672	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	113.340%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	36038	5.971	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	119.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	50589	5.417	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	108.400%	
Target Compounds						
5) Vinyl chloride	1.314	62	8579	0.752	ug/L	94
14) Carbon disulfide	2.301	76	2566	0.087	ug/L #	93
19) 1,1-Dichloroethane	3.195	63	2823	0.167	ug/L #	83
22) cis-1,2-Dichloroethene	3.918	96	2575	0.256	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032322\
Data File : VV025135.D
Acq On : 23 Mar 2022 16:20
Operator : SY/MD
Sample : N2083-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW5Z3

Quant Time: Mar 24 01:17:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 24 01:12:55 2022
Response via : Initial Calibration

