

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032323\
 Data File : VW030315.D
 Acq On : 23 Mar 2023 16:58
 Operator : SY/MD
 Sample : 02053-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EYF72

Quant Time: Mar 24 02:52:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 24 02:42:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	135913	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	134881	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	64013	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	39775	4.419	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.400%
7) Chloroethane-d5	1.536	69	40271	4.610	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.200%
11) 1,1-Dichloroethene-d2	2.063	63	68477	3.206	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.200%
20) 2-Butanone-d5	3.825	46	107407	46.164	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.320%
24) Chloroform-d	4.262	84	100836	4.628	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	4.953	65	51429	5.003	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.000%
32) Benzene-d6	4.970	84	177447	4.669	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
36) 1,2-Dichloropropane-d6	5.998	67	58886	4.949	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.252	98	148868	4.265	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.400%
43) trans-1,3-Dichloroprop...	7.564	79	19045	4.251	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.000%
46) 2-Hexanone-d5	8.037	63	86295	47.934	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.860%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	43508	4.639	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.800%
66) 1,2-Dichlorobenzene-d4	11.571	152	56496	4.938	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032323\
Data File : VV030315.D
Acq On : 23 Mar 2023 16:58
Operator : SY/MD
Sample : 02053-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EYF72

Quant Time: Mar 24 02:52:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 24 02:42:05 2023
Response via : Initial Calibration

