

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032420\
 Data File : VV015112.D
 Acq On : 24 Mar 2020 09:46
 Operator : SY/MD
 Sample : VV0324WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK91

Manual Integrations
 APPROVED

apatel
 3/25/2020 4:52:37 PM

Quant Time: Mar 25 01:28:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 25 01:10:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	789954	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	750996	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	349577	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	328628	45.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.08%
7) Chloroethane-d5	1.58	69	288172	46.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.32%
11) 1,1-Dichloroethene-d2	2.12	63	470203m	38.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.20%
21) 2-Butanone-d5	3.94	46	411511	86.56	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.56%
24) Chloroform-d	4.38	84	490513	44.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.94%
26) 1,2-Dichloroethane-d4	5.06	65	319037	45.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.86%
32) Benzene-d6	5.08	84	1038087	45.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.28%
36) 1,2-Dichloropropane-d6	6.10	67	328663	44.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.30%
41) Toluene-d8	7.34	98	995753	47.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.42%
43) trans-1,3-Dichloropropene-	7.64	79	172624	43.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.82%
47) 2-Hexanone-d5	8.11	63	326180	87.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.21%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	447636	42.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.76%
64) 1,2-Dichlorobenzene-d4	11.65	152	362946	46.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.00%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032420\
Data File : VV015112.D
Acq On : 24 Mar 2020 09:46
Operator : SY/MD
Sample : VV0324WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK91

Manual Integrations
APPROVED

apatel
3/25/2020 4:52:37 PM

Quant Time: Mar 25 01:28:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032320WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 25 01:10:42 2020
Response via : Initial Calibration

