

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032421\
 Data File : VW020729.D
 Acq On : 24 Mar 2021 12:09
 Operator : SY/MD
 Sample : M1783-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0G07

Quant Time: Mar 25 03:12:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 25 03:10:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	680501	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	663973	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	328355	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	221913	45.93	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	91.86%		
7) Chloroethane-d5	1.568	69	170155	46.59	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	93.18%		
11) 1,1-Dichloroethene-d2	2.108	63	336899	36.96	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	73.92%		
21) 2-Butanone-d5	3.889	46	309072	98.33	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	98.33%		
24) Chloroform-d	4.352	84	466117	48.54	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.08%		
26) 1,2-Dichloroethane-d4	5.034	65	313637	49.58	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.16%		
32) Benzene-d6	5.053	84	878023	49.61	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.22%		
36) 1,2-Dichloropropane-d6	6.072	67	271784	50.02	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	100.04%		
41) Toluene-d8	7.320	98	790237	47.97	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.94%		
43) trans-1,3-Dichloroprop...	7.625	79	134096	46.38	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	92.76%		
47) 2-Hexanone-d5	8.092	63	229785	100.87	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	100.87%		
57) 1,1,2,2-Tetrachloroeth...	10.220	84	361200	45.64	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.28%		
64) 1,2-Dichlorobenzene-d4	11.632	152	327002	51.31	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	102.62%		
Target Compounds						Qvalue
25) Chloroform	4.381	83	55571	6.18	ug/L	99
65) 1,2-Dichlorobenzene	11.651	146	8113	0.82	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032421\
Data File : VV020729.D
Acq On : 24 Mar 2021 12:09
Operator : SY/MD
Sample : M1783-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0G07

Quant Time: Mar 25 03:12:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 25 03:10:36 2021
Response via : Initial Calibration

