

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032422\
 Data File : VW025156.D
 Acq On : 24 Mar 2022 17:43
 Operator : SY/MD
 Sample : N2083-14
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW614

Quant Time: Mar 25 03:11:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 25 02:51:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	140436	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	135914	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62003	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	61089	5.510	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	110.200%	
7) Chloroethane-d5	1.568	69	48250	5.943	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	118.800%	
11) 1,1-Dichloroethene-d2	2.111	63	78459	4.378	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	87.600%	
20) 2-Butanone-d5	3.896	46	84538	63.091	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	126.180%	
24) Chloroform-d	4.352	84	104685	5.693	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	113.800%	
26) 1,2-Dichloroethane-d4	5.034	65	44479	5.786	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	115.800%	
32) Benzene-d6	5.050	84	221594	5.620	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	112.400%	
36) 1,2-Dichloropropane-d6	6.069	67	63418	5.594	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	111.800%	
41) Toluene-d8	7.317	98	192043	5.344	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.800%	
43) trans-1,3-Dichloroprop...	7.625	79	18864	4.773	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.400%	
46) 2-Hexanone-d5	8.088	63	71512	53.045	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	106.100%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	37819	5.670	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	113.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	64262	6.122	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	122.400%#	
Target Compounds						Qvalue
13) Acetone	2.185	43	5403	6.029	ug/L	82
25) Chloroform	4.375	83	6644	0.348	ug/L	93
38) Bromodichloromethane	6.523	83	1807	0.141	ug/L #	93
42) Toluene	7.397	91	11528	0.249	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032422\
Data File : VV025156.D
Acq On : 24 Mar 2022 17:43
Operator : SY/MD
Sample : N2083-14
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW614

Quant Time: Mar 25 03:11:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 25 02:51:12 2022
Response via : Initial Calibration

