

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032422\
 Data File : VW025163.D
 Acq On : 24 Mar 2022 20:27
 Operator : SY/MD
 Sample : N2100-08
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AG0

Quant Time: Mar 25 03:13:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 25 02:51:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	136264	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	132766	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58476	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	44608	4.147	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.000%	
7) Chloroethane-d5	1.571	69	42142	5.350	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.000%	
11) 1,1-Dichloroethene-d2	2.111	63	61893	3.560	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	71.200%	
20) 2-Butanone-d5	3.899	46	102907	79.152	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	158.300%#	
24) Chloroform-d	4.352	84	101010	5.661	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	113.200%	
26) 1,2-Dichloroethane-d4	5.034	65	45219	6.062	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	121.200%	
32) Benzene-d6	5.053	84	198395	5.151	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.000%	
36) 1,2-Dichloropropane-d6	6.072	67	62774	5.668	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	113.400%	
41) Toluene-d8	7.316	98	166648	4.748	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.000%	
43) trans-1,3-Dichloroprop...	7.625	79	17661	4.574	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.400%	
46) 2-Hexanone-d5	8.088	63	84412	64.099	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	128.200%	
56) 1,1,2,2-Tetrachloroeth...	10.213	84	42222	6.480	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	129.600%#	
66) 1,2-Dichlorobenzene-d4	11.622	152	61439	6.206	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	124.200%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032422\
Data File : VV025163.D
Acq On : 24 Mar 2022 20:27
Operator : SY/MD
Sample : N2100-08
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AG0

Quant Time: Mar 25 03:13:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 25 02:51:12 2022
Response via : Initial Calibration

