

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032422\
 Data File : VW025164.D
 Acq On : 24 Mar 2022 20:50
 Operator : SY/MD
 Sample : N2100-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AG1

Quant Time: Mar 25 03:13:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 25 02:51:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	137189	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	132875	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58752	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	43530	4.019	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	80.400%	
7) Chloroethane-d5	1.571	69	41333	5.211	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.200%	
11) 1,1-Dichloroethene-d2	2.111	63	59677	3.409	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.200%	
20) 2-Butanone-d5	3.895	46	103527	79.092	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	158.180%#	
24) Chloroform-d	4.352	84	98651	5.492	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.800%	
26) 1,2-Dichloroethane-d4	5.034	65	44302	5.899	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	118.000%	
32) Benzene-d6	5.053	84	193759	5.027	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
36) 1,2-Dichloropropane-d6	6.069	67	61164	5.518	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	110.400%	
41) Toluene-d8	7.317	98	160761	4.576	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.600%	
43) trans-1,3-Dichloroprop...	7.625	79	17639	4.565	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.200%	
46) 2-Hexanone-d5	8.088	63	82508	62.601	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	125.200%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42015	6.443	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	128.800%#	
66) 1,2-Dichlorobenzene-d4	11.622	152	59579	5.990	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	119.800%	
Target Compounds						
9) Trichlorofluoromethane	1.757	101	2500	0.152	ug/L	99
17) Methyl tert-butyl Ether	2.780	73	2079	0.109	ug/L #	86
25) Chloroform	4.378	83	4503	0.242	ug/L	91
47) Tetrachloroethene	7.982	164	1585	0.196	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032422\
Data File : VV025164.D
Acq On : 24 Mar 2022 20:50
Operator : SY/MD
Sample : N2100-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AG1

Quant Time: Mar 25 03:13:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR031522WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 25 02:51:12 2022
Response via : Initial Calibration

