

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV032425\
 Data File : VV038735.D
 Acq On : 24 Mar 2025 13:05
 Operator : SY/MD
 Sample : VSTD00242
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002242

Quant Time: Mar 24 13:36:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM032425.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Mar 24 13:19:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5762	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5520	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2498	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	10167	1.91	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 382.00%#			
4) 1,2-Dichloroethane-d4	4.945	65	5301	1.30	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery = 260.00%#			
7) 1,2-Dichloropropane-d6	5.995	67	6570	1.27	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 254.00%#			
8) Toluene-d8	7.238	98	12969	1.07	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery = 214.00%#			
10) 1,1,2,2-Tetrachloroeth...	10.145	84	5410	1.35	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 270.00%#			
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	19338	2.83	ug/L	99
6) Trichloroethene	5.828	95	10591	2.26	ug/L	98
9) 1,2-Dibromoethane	8.289	107	7437	2.46	ug/L #	98
12) 1,2,3-Trichloropropane	10.197	75	6427	2.47	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.356	75	1300	2.91	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV032425\
Data File : VV038735.D
Acq On : 24 Mar 2025 13:05
Operator : SY/MD
Sample : VSTD00242
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD00242

Quant Time: Mar 24 13:36:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM032425.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Mar 24 13:19:44 2025
Response via : Initial Calibration

