

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV032425\
 Data File : VV038736.D
 Acq On : 24 Mar 2025 13:35
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Quant Time: Mar 24 15:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM032425.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Mar 24 15:12:15 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							
1) 1,4-Difluorobenzene	5.566	114	5662	0.50	ug/L	0.00	
5) Chlorobenzene-d5	8.800	117	5381	0.50	ug/L	0.00	
11) 1,4-Dichlorobenzene-d4	11.191	152	2470	0.50	ug/L	0.00	
System Monitoring Compounds							
2) Vinyl Chloride-d3	1.280	65	2621	0.49	ug/L	0.00	
4) 1,2-Dichloroethane-d4	4.969	65	1349	0.50	ug/L	0.00	
7) 1,2-Dichloropropane-d6	5.995	67	1556	0.47	ug/L	0.00	
8) Toluene-d8	7.253	98	3092	0.50	ug/L	0.00	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1348	0.48	ug/L	0.00	
Target Compounds							
3) Vinyl chloride	1.280	62	4976	0.52	ug/L		Qvalue 93
6) Trichloroethene	5.852	95	2664	0.55	ug/L		99
9) 1,2-Dibromoethane	8.305	107	1803	0.52	ug/L #		99
12) 1,2,3-Trichloropropane	10.197	75	1659	0.55	ug/L		98
13) 1,2-Dibromo-3-chloropr...	12.357	75	344	0.51	ug/L #		80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Client Sample :
 Instrument :

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV032425\
 Data File : VV038736.D
 Acq On : 24 Mar 2025 13:35
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Quant Time: Mar 24 15:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM032425.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Mar 24 15:12:15 2025
 Response via : Initial Calibration

