

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032519\
 Data File : VV009862.D
 Acq On : 25 Mar 2019 15:44
 Operator : SY/MD
 Sample : VV0325WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK78

Quant Time: Mar 26 07:25:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 07:13:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	182512	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	176310	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	72283	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93981	50.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.86%
7) Chloroethane-d5	1.57	69	104389	49.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.56%
11) 1,1-Dichloroethene-d2	2.13	63	199540	36.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.64%
21) 2-Butanone-d5	3.93	46	76787	92.32	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.32%
24) Chloroform-d	4.40	84	114899	44.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.68%
26) 1,2-Dichloroethane-d4	5.08	65	82870	47.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.68%
32) Benzene-d6	5.10	84	258698	54.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.12%
36) 1,2-Dichloropropane-d6	6.12	67	77186	52.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.56%
41) Toluene-d8	7.36	98	233484	49.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.64%
43) trans-1,3-Dichloropropene-	7.66	79	29748	44.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.22%
47) 2-Hexanone-d5	8.13	63	56057	92.63	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.63%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	103811	43.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.66%
64) 1,2-Dichlorobenzene-d4	11.67	152	78119	46.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.30%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032519\
 Data File : VV009862.D
 Acq On : 25 Mar 2019 15:44
 Operator : SY/MD
 Sample : VV0325WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 VBLK78

Quant Time: Mar 26 07:25:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 07:13:38 2019
 Response via : Initial Calibration

