

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032519\
 Data File : VV009884.D
 Acq On : 26 Mar 2019 00:47
 Operator : SY/MD
 Sample : K2063-11
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB6S5

Quant Time: Mar 26 09:02:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 07:13:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	163612	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	157719	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61903	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	85369	51.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.20%
7) Chloroethane-d5	1.56	69	94578	50.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.62%
11) 1,1-Dichloroethene-d2	2.13	63	182238	37.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.00%
21) 2-Butanone-d5	3.93	46	78653	105.48	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.48%
24) Chloroform-d	4.40	84	110614	47.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.24%
26) 1,2-Dichloroethane-d4	5.08	65	77085	49.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.28%
32) Benzene-d6	5.10	84	234532	54.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.58%
36) 1,2-Dichloropropane-d6	6.12	67	71650	54.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.50%
41) Toluene-d8	7.36	98	207163	49.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.84%
43) trans-1,3-Dichloropropene-	7.66	79	25867	42.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.76%
47) 2-Hexanone-d5	8.14	63	56463	104.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.30%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	94627	44.15	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.30%
64) 1,2-Dichlorobenzene-d4	11.67	152	71788	50.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.12%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032519\
Data File : VV009884.D
Acq On : 26 Mar 2019 00:47
Operator : SY/MD
Sample : K2063-11
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB6S5

Quant Time: Mar 26 09:02:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 26 07:13:38 2019
Response via : Initial Calibration

