

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032519\
 Data File : VV009885.D
 Acq On : 26 Mar 2019 01:11
 Operator : SY/MD
 Sample : K2084-18
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF7Z6

Quant Time: Mar 26 09:07:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 07:13:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	169479	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	162687	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	63224	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83441	48.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.44%
7) Chloroethane-d5	1.56	69	93161	47.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.68%
11) 1,1-Dichloroethene-d2	2.13	63	182232	35.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.44%
21) 2-Butanone-d5	3.93	46	73400	95.03	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.03%
24) Chloroform-d	4.40	84	111607	46.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.76%
26) 1,2-Dichloroethane-d4	5.08	65	78339	48.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.40%
32) Benzene-d6	5.10	84	236586	53.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.16%
36) 1,2-Dichloropropane-d6	6.12	67	71275	52.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.64%
41) Toluene-d8	7.36	98	205384	47.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	7.67	79	26035	41.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.68%
47) 2-Hexanone-d5	8.13	63	51631	92.46	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.46%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	97024	43.89	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.78%
64) 1,2-Dichlorobenzene-d4	11.67	152	71885	49.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032519\
Data File : VV009885.D
Acq On : 26 Mar 2019 01:11
Operator : SY/MD
Sample : K2084-18
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF7Z6

Quant Time: Mar 26 09:07:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 26 07:13:38 2019
Response via : Initial Calibration

