

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032519\
 Data File : VV009881.D
 Acq On : 25 Mar 2019 23:35
 Operator : SY/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK37

Quant Time: Mar 26 08:46:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 07:13:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	172703	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	168273	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	67272	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83628	47.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.84%
7) Chloroethane-d5	1.56	69	92716	46.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.46%
11) 1,1-Dichloroethene-d2	2.13	63	178806	34.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.80%
21) 2-Butanone-d5	3.93	46	78911	100.26	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.26%
24) Chloroform-d	4.40	84	109369	44.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.20%
26) 1,2-Dichloroethane-d4	5.08	65	78011	47.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.18%
32) Benzene-d6	5.10	84	235537	51.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.14%
36) 1,2-Dichloropropane-d6	6.12	67	73390	52.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.18%
41) Toluene-d8	7.36	98	211026	47.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.36%
43) trans-1,3-Dichloropropene-	7.67	79	26820	41.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.34%
47) 2-Hexanone-d5	8.14	63	54234	93.90	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.90%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	98651	43.14	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.28%
64) 1,2-Dichlorobenzene-d4	11.67	152	71001	45.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.12%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032519\
Data File : VV009881.D
Acq On : 25 Mar 2019 23:35
Operator : SY/MD
Sample : IBLK
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK37

Quant Time: Mar 26 08:46:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 26 07:13:38 2019
Response via : Initial Calibration

