

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032519\
 Data File : VV009886.D
 Acq On : 26 Mar 2019 01:36
 Operator : SY/MD
 Sample : K2063-17
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DB6S6

Quant Time: Mar 26 09:10:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 07:13:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	169998	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	162676	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61296	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82060	47.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.56%
7) Chloroethane-d5	1.56	69	91932	47.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.14%
11) 1,1-Dichloroethene-d2	2.13	63	177188	34.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.26%
21) 2-Butanone-d5	3.93	46	76076	98.19	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.19%
24) Chloroform-d	4.40	84	107525	44.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.10%
26) 1,2-Dichloroethane-d4	5.08	65	75183	46.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.20%
32) Benzene-d6	5.10	84	228628	51.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.56%
36) 1,2-Dichloropropane-d6	6.12	67	71487	52.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.96%
41) Toluene-d8	7.36	98	204338	47.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.52%
43) trans-1,3-Dichloropropene-	7.66	79	25132	40.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.78%
47) 2-Hexanone-d5	8.13	63	51880	92.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.92%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	94333	42.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.34%
64) 1,2-Dichlorobenzene-d4	11.68	152	70096	49.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.72%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032519\
Data File : VV009886.D
Acq On : 26 Mar 2019 01:36
Operator : SY/MD
Sample : K2063-17
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DB6S6

Quant Time: Mar 26 09:10:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 26 07:13:38 2019
Response via : Initial Calibration

