

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032620\
 Data File : VV015190.D
 Acq On : 27 Mar 2020 01:26
 Operator : SY/MD
 Sample : L1976-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTL1

Quant Time: Mar 27 06:36:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 27 03:08:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	559947	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	530497	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	214066	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	146933	48.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.82%
7) Chloroethane-d5	1.58	69	149194	49.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.52%
11) 1,1-Dichloroethene-d2	2.12	63	226635	33.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.16%
21) 2-Butanone-d5	3.94	46	275206	92.00	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.00%
24) Chloroform-d	4.38	84	304695	46.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.56%
26) 1,2-Dichloroethane-d4	5.06	65	212204	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.64%
32) Benzene-d6	5.08	84	582756	53.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.80%
36) 1,2-Dichloropropane-d6	6.10	67	198999	50.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.12%
41) Toluene-d8	7.34	98	535789	54.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.36%
43) trans-1,3-Dichloropropene-	7.64	79	71406	33.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.62%
47) 2-Hexanone-d5	8.12	63	198417	86.21	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.21%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	300266	44.92	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.84%
64) 1,2-Dichlorobenzene-d4	11.65	152	202034	50.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.10%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	85911	15.631	ug/L	95
20) cis-1,2-Dichloroethene	3.95	96	7292	1.668	ug/L #	95
25) Chloroform	4.41	83	10977	1.470	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032620\
 Data File : VV015190.D
 Acq On : 27 Mar 2020 01:26
 Operator : SY/MD
 Sample : L1976-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 ETTL1

Quant Time: Mar 27 06:36:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032620WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 27 03:08:32 2020
 Response via : Initial Calibration

