

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032621\
 Data File : VW020746.D
 Acq On : 26 Mar 2021 11:03
 Operator : SY/MD
 Sample : VW0326MBL01
 Misc : 5.0g/5mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK79

Quant Time: Mar 27 06:51:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 27 06:50:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	688805	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	671529	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	329069	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.307	65	222300	45.46	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.92%
7) Chloroethane-d5	1.568	69	176130	47.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.28%
11) 1,1-Dichloroethene-d2	2.108	63	342997	37.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.36%
21) 2-Butanone-d5	3.886	46	289842	91.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.10%
24) Chloroform-d	4.349	84	472761	48.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.28%
26) 1,2-Dichloroethane-d4	5.031	65	321194	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
32) Benzene-d6	5.050	84	897555	50.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.28%
36) 1,2-Dichloropropane-d6	6.072	67	272261	49.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.08%
41) Toluene-d8	7.317	98	813202	48.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.62%
43) trans-1,3-Dichloroprop...	7.622	79	137752	47.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.22%
47) 2-Hexanone-d5	8.092	63	212155	92.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.08%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	370094	46.23	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.46%
64) 1,2-Dichlorobenzene-d4	11.632	152	336762	52.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032621\
Data File : VV020746.D
Acq On : 26 Mar 2021 11:03
Operator : SY/MD
Sample : VV0326MBL01
Misc : 5.0g/5mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK79

Quant Time: Mar 27 06:51:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 27 06:50:46 2021
Response via : Initial Calibration

