

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032621\
 Data File : VW020750.D
 Acq On : 26 Mar 2021 12:37
 Operator : SY/MD
 Sample : M1816-05 10X
 Misc : 4.34g/5mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3L3

Quant Time: Mar 27 06:52:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 27 06:50:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	570568	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	556721	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	272266	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	192039	47.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.82%
7) Chloroethane-d5	1.571	69	166693	54.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.86%
11) 1,1-Dichloroethene-d2	2.108	63	314888	41.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.40%
21) 2-Butanone-d5	3.892	46	327958	124.44	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.44%
24) Chloroform-d	4.352	84	496633	61.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	123.36%
26) 1,2-Dichloroethane-d4	5.037	65	336860	63.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	127.02%#
32) Benzene-d6	5.053	84	894949	60.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	120.62%
36) 1,2-Dichloropropane-d6	6.072	67	286376	62.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	125.72%#
41) Toluene-d8	7.320	98	791884	57.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.66%
43) trans-1,3-Dichloroprop...	7.625	79	143574	59.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	118.46%
47) 2-Hexanone-d5	8.092	63	247039	129.33	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	129.33%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	407128	61.35	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	122.70%#
64) 1,2-Dichlorobenzene-d4	11.632	152	352211	66.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	133.30%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032621\
Data File : VV020750.D
Acq On : 26 Mar 2021 12:37
Operator : SY/MD
Sample : M1816-05 10X
Misc : 4.34g/5mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3L3

Quant Time: Mar 27 06:52:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 27 06:50:46 2021
Response via : Initial Calibration

