

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032624\
 Data File : VV034830.D
 Acq On : 26 Mar 2024 16:12
 Operator : SY/MD
 Sample : VV0326WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK254

Quant Time: Mar 27 01:36:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 27 01:34:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	382374	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	358213	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	165350	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	160168	43.373	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.740%
7) Chloroethane-d5	1.532	69	138454	43.167	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.340%
11) 1,1-Dichloroethene-d2	2.060	65	71361	44.084	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.160%
21) 2-Butanone-d5	3.783	46	201251	91.890	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.890%
24) Chloroform-d	4.246	84	268900	43.310	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.620%
26) 1,2-Dichloroethane-d4	4.941	65	176759	44.841	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.680%
32) Benzene-d6	4.960	84	554510	46.084	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.160%
36) 1,2-Dichloropropane-d6	5.989	67	178395	46.689	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.380%
41) Toluene-d8	7.243	98	486518	44.897	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	7.551	79	78771	42.406	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.820%
47) 2-Hexanone-d5	8.024	63	129722	89.796	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.800%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	232339	44.157	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.320%
66) 1,2-Dichlorobenzene-d4	11.558	152	173178	46.648	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032624\
Data File : VV034830.D
Acq On : 26 Mar 2024 16:12
Operator : SY/MD
Sample : VV0326WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK254

Quant Time: Mar 27 01:36:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 27 01:34:52 2024
Response via : Initial Calibration

