

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032624\
 Data File : VW034831.D
 Acq On : 26 Mar 2024 16:41
 Operator : SY/MD
 Sample : P1851-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0LX1

Quant Time: Mar 27 01:36:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 27 01:34:52 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.535	114	144470	50.000	ug/L	0.00
28) Chlorobenzene-d5		8.783	117	138845	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.185	152	64796	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	136452	97.799	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	195.600%#	
7) Chloroethane-d5		1.532	69	115557	95.356	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	190.720%#	
11) 1,1-Dichloroethene-d2		2.060	65	59910	97.957	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	195.920%#	
21) 2-Butanone-d5		3.786	46	150145	181.449	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	181.450%#	
24) Chloroform-d		4.246	84	225292	96.040	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	192.080%#	
26) 1,2-Dichloroethane-d4		4.937	65	149822	100.596	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	201.200%#	
32) Benzene-d6		4.957	84	464442	99.583	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	199.160%#	
36) 1,2-Dichloropropane-d6		5.986	67	150470	101.600	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	203.200%#	
41) Toluene-d8		7.243	98	406426	96.763	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	193.520%#	
43) trans-1,3-Dichloroprop...		7.551	79	66273	92.048	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	184.100%#	
47) 2-Hexanone-d5		8.024	63	97560	174.232	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	174.230%#	
56) 1,1,2,2-Tetrachloroeth...		10.149	84	196417	96.308	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	192.620%#	
66) 1,2-Dichlorobenzene-d4		11.558	152	142832	98.179	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	196.360%#	
Target Compounds							
13) Acetone		2.118	43	3069	4.040	ug/L	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032624\
Data File : VV034831.D
Acq On : 26 Mar 2024 16:41
Operator : SY/MD
Sample : P1851-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0LX1

Quant Time: Mar 27 01:36:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 27 01:34:52 2024
Response via : Initial Calibration

