

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032624\
 Data File : VW034842.D
 Acq On : 26 Mar 2024 21:04
 Operator : SY/MD
 Sample : P1852-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0LY3

Quant Time: Mar 27 01:39:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 27 01:34:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	165571	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	164853	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	72087	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	121503	75.987	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	151.980%#
7) Chloroethane-d5	1.532	69	105914	76.261	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	152.520%#
11) 1,1-Dichloroethene-d2	2.059	65	54296	77.463	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	154.920%#
21) 2-Butanone-d5	3.789	46	131250	138.400	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	138.400%#
24) Chloroform-d	4.249	84	208288	77.475	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	154.960%#
26) 1,2-Dichloroethane-d4	4.940	65	140071	82.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	164.120%#
32) Benzene-d6	4.960	84	428224	77.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	154.660%#
36) 1,2-Dichloropropane-d6	5.989	67	137562	78.230	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	156.460%#
41) Toluene-d8	7.243	98	366573	73.506	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	147.020%#
43) trans-1,3-Dichloroprop...	7.554	79	59071	69.101	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	138.200%#
47) 2-Hexanone-d5	8.024	63	85085	127.980	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	127.980%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	180835	74.679	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	149.360%#
66) 1,2-Dichlorobenzene-d4	11.561	152	132708	81.994	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	163.980%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032624\
Data File : VV034842.D
Acq On : 26 Mar 2024 21:04
Operator : SY/MD
Sample : P1852-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0LY3

Quant Time: Mar 27 01:39:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 27 01:34:52 2024
Response via : Initial Calibration

