

Data File : VV009952.D
Acq On : 28 Mar 2019 02:58
Operator : SY/MD
Sample : K2101-15
Misc : 5.96G/10mL/MSVOA_V/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF0E6

Quant Time: Mar 29 13:39:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM032619S.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 07:17:22 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	169594	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	168147	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	66810	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	72993	21.21	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.84%
7) Chloroethane-d5	1.57	69	84197	22.65	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.60%
10) 1,1-Dichloroethene-d2	2.12	63	161784	17.19	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.76%
20) 2-Butanone-d5	3.93	46	25903	40.39	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.78%
24) Chloroform-d	4.40	84	90212	20.88	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.52%
26) 1,2-Dichloroethane-d4	5.08	65	59695	22.96	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.84%
29) Benzene-d6	5.10	84	213404	24.17	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.68%
33) 1,2-Dichloropropane-d6	6.12	67	65783	25.46	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.84%
37) Toluene-d8	7.36	98	188733	22.58	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.32%
38) trans-1,3-Dichloropropene-	7.66	79	19988	18.91	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.64%
39) 2-Hexanone-d5	8.13	63	17574	36.44	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.88%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	48556	18.98	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	75.92%
61) 1,2-Dichlorobenzene-d4	11.67	152	59035	23.21	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.84%

Target Compounds

					Qvalue
5) Vinyl chloride	1.32	62	5900	1.495 ug/L	# 70
16) Methylene chloride	2.53	84	6879	2.463 ug/L	95
22) cis-1,2-Dichloroethene	3.96	96	141509	57.673 ug/L	95
35) Trichloroethene	5.96	95	230639	94.696 ug/L	98
47) Tetrachloroethene	8.02	164	996886	518.690 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV009952.D
Acq On : 28 Mar 2019 02:58
Operator : SY/MD
Sample : K2101-15
Misc : 5.96G/10mL/MSVOA_V/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF0E6

Quant Time: Mar 29 13:39:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM032619S.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 07:17:22 2019
Response via : Initial Calibration

