

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032720\
 Data File : VV015230.D
 Acq On : 28 Mar 2020 07:10
 Operator : SY/MD
 Sample : L1976-19
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTK4

Quant Time: Mar 28 08:22:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 28 05:21:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	651580	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	609571	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	258178	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	138127	52.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.50%
7) Chloroethane-d5	1.58	69	138715	48.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.90%
11) 1,1-Dichloroethene-d2	2.12	63	218561	33.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.12%
21) 2-Butanone-d5	3.94	46	219383	70.05	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.05%
24) Chloroform-d	4.38	84	287901	41.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.50%
26) 1,2-Dichloroethane-d4	5.06	65	189369	42.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.28%
32) Benzene-d6	5.07	84	538916	50.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.42%
36) 1,2-Dichloropropane-d6	6.09	67	182495	45.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.80%
41) Toluene-d8	7.34	98	482585	51.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.04%
43) trans-1,3-Dichloropropene-	7.64	79	87782	40.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.72%
47) 2-Hexanone-d5	8.12	63	162692	70.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	70.12%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	263540	37.22	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	74.44%
64) 1,2-Dichlorobenzene-d4	11.65	152	183925	41.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.52%

Target Compounds

					Ovalue
25) Chloroform	4.40	83	8215	0.982 ug/L	96
51) Chlorobenzene	8.90	112	39991	3.252 ug/L	99
62) 1,3-Dichlorobenzene	11.21	146	4419	0.550 ug/L	95
63) 1,4-Dichlorobenzene	11.30	146	33088	4.097 ug/L	95
65) 1,2-Dichlorobenzene	11.67	146	25312	3.072 ug/L	95
68) 1,2,4-trichlorobenzene	13.29	180	10092	2.127 ug/L	96
70) 1,2,3-Trichlorobenzene	13.77	180	2328	0.478 ug/L #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV032720\
Data File : VV015230.D
Acq On : 28 Mar 2020 07:10
Operator : SY/MD
Sample : L1976-19
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTK4

Quant Time: Mar 28 08:22:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032720WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 28 05:21:13 2020
Response via : Initial Calibration

