

Quantitation Report (Qedit)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032720\
 Data File : VV015246.D
 Acq On : 28 Mar 2020 15:08
 Operator : SY/MD
 Sample : L1975-17 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTJ8

Manual Integrations
 APPROVED

apatel
 3/31/2020 9:24:27 AM

Quant Time: Mar 30 02:15:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 30 02:13:00 2020
 Response via : Initial Calibration



(11) 1,1-Dichloroethene-d2 (S)

2.119min (+0.000) 33.50ug/L m >

response 211581

Ion	Exp%	Act%
63.00	100	100
98.00	79.40	0.08#
65.00	19.10	0.02#
0.00	0.00	0.00

SY
 03/31/2020

Quantitation Report (QT Reviewed)

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032720\
 Data File : VV015246.D
 Acq On : 28 Mar 2020 15:08
 Operator : SY/MD
 Sample : L1975-17 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTJ8

Manual Integrations
 APPROVED

apatel
 3/31/2020 9:24:27 AM

Quant Time: Mar 30 03:21:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 30 02:13:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	622517	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	585695	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	253019	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.32	65	124147	49.15	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	98.30%		
7) Chloroethane-d5	1.58	69	128375	47.41	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	94.82%		
11) 1,1-Dichloroethene-d2	2.12	63	211581m	33.50	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	67.00%		
21) 2-Butanone-d5	3.94	46	220403	73.66	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	73.66%		
24) Chloroform-d	4.38	84	271837	41.26	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.52%		
26) 1,2-Dichloroethane-d4	5.06	65	185050	43.10	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.20%		
32) Benzene-d6	5.08	84	502122	48.69	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.38%		
36) 1,2-Dichloropropane-d6	6.10	67	172519	44.66	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.32%		
41) Toluene-d8	7.34	98	458037	50.40	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	100.80%		
43) trans-1,3-Dichloropropene-	7.64	79	87428	42.35	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	84.70%		
47) 2-Hexanone-d5	8.12	63	169415	75.99	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	75.99%		
57) 1,1,2,2-Tetrachloroethane-	10.24	84	256286	37.67	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	75.34%		
64) 1,2-Dichlorobenzene-d4	11.65	152	180156	41.74	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	83.48%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
25) Chloroform	4.41	83	8523	1.067	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

SY
 03/31/2020

Quantitation Report (QT Reviewed)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032720\
 Data File : VV015246.D
 Acq On : 28 Mar 2020 15:08
 Operator : SY/MD
 Sample : L1975-17 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTJ8

Manual Integrations
 APPROVED

apatel
 3/31/2020 9:24:27 AM

Quant Time: Mar 30 03:21:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 30 02:13:00 2020
 Response via : Initial Calibration



Quantitation Report (Qedit)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV032720\
 Data File : VV015246.D
 Acq On : 28 Mar 2020 15:08
 Operator : SY/MD
 Sample : L1975-17 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTJ8

Manual Integrations
 APPROVED

apatel
 3/31/2020 9:24:27 AM

Quant Time: Mar 30 02:15:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM032720WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 30 02:13:00 2020
 Response via : Initial Calibration



(11) 1,1-Dichloroethene-d2 (S)

2.238min (+0.119) 0.04ug/L

response 245

Ion	Exp%	Act%
63.00	100	100
98.00	79.40	72.65
65.00	19.10	19.18
0.00	0.00	0.00