

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV033020\
 Data File : VV015294.D
 Acq On : 31 Mar 2020 00:57
 Operator : SY/MD
 Sample : VV0330WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Mar 31 06:34:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 31 06:20:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	595775	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	575612	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	238793	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	193784	44.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.22%
7) Chloroethane-d5	1.58	69	184188	45.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.12%
11) 1,1-Dichloroethene-d2	2.12	63	311050m	33.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.40%
21) 2-Butanone-d5	3.93	46	274923	96.78	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.78%
24) Chloroform-d	4.38	84	362304	46.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.78%
26) 1,2-Dichloroethane-d4	5.06	65	251821	49.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.66%
32) Benzene-d6	5.08	84	706334	47.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.02%
36) 1,2-Dichloropropane-d6	6.10	67	226468	48.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.40%
41) Toluene-d8	7.34	98	670281	46.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.22%
43) trans-1,3-Dichloropropene-	7.64	79	112851	44.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.20%
47) 2-Hexanone-d5	8.12	63	233291	105.83	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.83%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	329587	47.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.06%
64) 1,2-Dichlorobenzene-d4	11.65	152	230768	48.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV033020\
Data File : VV015294.D
Acq On : 31 Mar 2020 00:57
Operator : SY/MD
Sample : VV0330WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Mar 31 06:34:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 31 06:20:52 2020
Response via : Initial Calibration

