

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV033020\
 Data File : VV015272.D
 Acq On : 30 Mar 2020 15:14
 Operator : SY/MD
 Sample : VV0330WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK97

Quant Time: Mar 31 02:28:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 31 02:12:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	700193	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	657116	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	286576	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	213898	41.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.80%
7) Chloroethane-d5	1.58	69	206167	43.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.78%
11) 1,1-Dichloroethene-d2	2.12	63	371737	33.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.52%
21) 2-Butanone-d5	3.94	46	302566	90.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.62%
24) Chloroform-d	4.38	84	386691	42.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.26%
26) 1,2-Dichloroethane-d4	5.06	65	266936	44.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.00%
32) Benzene-d6	5.08	84	779338	45.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.86%
36) 1,2-Dichloropropane-d6	6.10	67	245337	45.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.48%
41) Toluene-d8	7.34	98	739174	44.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.08%
43) trans-1,3-Dichloropropene-	7.64	79	125035	42.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.60%
47) 2-Hexanone-d5	8.12	63	236343	93.91	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.91%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	352526	44.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.06%
64) 1,2-Dichlorobenzene-d4	11.65	152	250524	43.92	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.84%

Target Compounds					Ovalue
25) Chloroform	4.41	83	11353	1.314 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV033020\
Data File : VV015272.D
Acq On : 30 Mar 2020 15:14
Operator : SY/MD
Sample : VV0330WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK97

Quant Time: Mar 31 02:28:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 31 02:12:40 2020
Response via : Initial Calibration

