

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV033020\
 Data File : VV015283.D
 Acq On : 30 Mar 2020 19:43
 Operator : SY/MD
 Sample : L1975-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETTH5

Manual Integrations
 APPROVED

MMDadoda
 3/31/2020 4:22:16 PM

Quant Time: Mar 31 05:03:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 31 02:12:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	591420	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	570147	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	254405	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	110516	25.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	51.26%#
7) Chloroethane-d5	1.58	69	119841	29.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	59.72%#
11) 1,1-Dichloroethene-d2	2.12	63	179544m	19.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	38.62%#
21) 2-Butanone-d5	3.93	46	257482	91.30	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.30%
24) Chloroform-d	4.38	84	264984	34.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.36%#
26) 1,2-Dichloroethane-d4	5.06	65	177836	35.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.20%
32) Benzene-d6	5.08	84	475743	31.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	63.92%#
36) 1,2-Dichloropropane-d6	6.09	67	163091	35.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	70.08%
41) Toluene-d8	7.34	98	436632	30.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	60.64%#
43) trans-1,3-Dichloropropene-	7.64	79	72346	28.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.08%#
47) 2-Hexanone-d5	8.11	63	222423	101.87	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.87%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	265378	38.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	178477	35.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	70.50%#

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	299396	60.061	ug/L	99
13) Acetone	2.21	43	45926	23.865	ug/L	81
19) 1,1-Dichloroethane	3.22	63	6469	0.947	ug/L	93
20) cis-1,2-Dichloroethene	3.95	96	7449	1.808	ug/L #	92
22) 2-Butanone	4.01	43	106910	38.110	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV033020\
Data File : VV015283.D
Acq On : 30 Mar 2020 19:43
Operator : SY/MD
Sample : L1975-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETTH5

Manual Integrations
APPROVED

MMDadoda
3/31/2020 4:22:16 PM

Quant Time: Mar 31 05:03:26 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 31 02:12:40 2020
Response via : Initial Calibration

