

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
 Data File : VV030354.D
 Acq On : 30 Mar 2023 13:09
 Operator : SY/MD
 Sample : 02122-04 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB9A6

Quant Time: Mar 31 05:40:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 31 05:37:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	172232	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	167131	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	80857	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	32004	2.806	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	56.200%	
7) Chloroethane-d5	1.539	69	36550	3.302	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	66.000%	
11) 1,1-Dichloroethene-d2	2.066	63	57551	2.126	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	42.600%#	
20) 2-Butanone-d5	3.818	46	110804	37.581	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	75.160%	
24) Chloroform-d	4.262	84	96819	3.507	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	70.200%	
26) 1,2-Dichloroethane-d4	4.957	65	48801	3.746	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.000%	
32) Benzene-d6	4.969	84	164985	3.503	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	70.000%	
36) 1,2-Dichloropropane-d6	5.998	67	56442	3.828	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	76.600%	
41) Toluene-d8	7.252	98	137124	3.171	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	63.400%#	
43) trans-1,3-Dichloroprop...	7.564	79	17864	3.218	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	64.400%	
46) 2-Hexanone-d5	8.037	63	82906	37.165	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	74.320%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	41428	3.565	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	71.200%	
66) 1,2-Dichlorobenzene-d4	11.570	152	55778	3.860	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	77.200%#	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.455	84	8159	0.479	ug/L	93
33) Benzene	5.021	78	16619	0.280	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
Data File : VV030354.D
Acq On : 30 Mar 2023 13:09
Operator : SY/MD
Sample : 02122-04 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CB9A6

Quant Time: Mar 31 05:40:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 31 05:37:21 2023
Response via : Initial Calibration

