

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW033023\
 Data File : VW030359.D
 Acq On : 30 Mar 2023 15:09
 Operator : SY/MD
 Sample : 02132-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMA9

Quant Time: Mar 31 05:41:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 31 05:37:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	129263	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	127258	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	59980	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	34242	4.000	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.000%
7) Chloroethane-d5	1.536	69	39957	4.809	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.200%
11) 1,1-Dichloroethene-d2	2.066	63	61056	3.005	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.200%
20) 2-Butanone-d5	3.818	46	121332	54.832	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.660%
24) Chloroform-d	4.259	84	103111	4.976	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
26) 1,2-Dichloroethane-d4	4.953	65	52165	5.336	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	4.970	84	177871	4.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	5.998	67	58486	5.209	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.252	98	146667	4.454	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	7.564	79	19572	4.630	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.600%
46) 2-Hexanone-d5	8.034	63	87300	51.397	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.800%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	42959	4.855	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.000%
66) 1,2-Dichlorobenzene-d4	11.571	152	59443	5.545	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
Data File : VV030359.D
Acq On : 30 Mar 2023 15:09
Operator : SY/MD
Sample : 02132-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMA9

Quant Time: Mar 31 05:41:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 31 05:37:21 2023
Response via : Initial Calibration

