

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW033023\
 Data File : VW030360.D
 Acq On : 30 Mar 2023 15:34
 Operator : SY/MD
 Sample : 02132-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMB7

Quant Time: Mar 31 05:41:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 31 05:37:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	126561	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	122224	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	58321	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	33990	4.055	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.200%
7) Chloroethane-d5	1.532	69	35788	4.400	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	2.056	63	61577	3.096	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.000%
20) 2-Butanone-d5	3.899	46	120294	55.523	ug/L	0.08
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.040%
24) Chloroform-d	4.256	84	99996	4.929	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
26) 1,2-Dichloroethane-d4	4.953	65	50157	5.240	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	4.960	84	169364	4.917	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
36) 1,2-Dichloropropane-d6	5.995	67	55684	5.164	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.200%
41) Toluene-d8	7.249	98	133935	4.235	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	7.564	79	17541	4.321	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%
46) 2-Hexanone-d5	8.050	63	77651	47.599	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.200%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	42354	4.984	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	11.570	152	59889	5.745	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
Data File : VV030360.D
Acq On : 30 Mar 2023 15:34
Operator : SY/MD
Sample : 02132-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMB7

Quant Time: Mar 31 05:41:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 31 05:37:21 2023
Response via : Initial Calibration

