

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW033023\
 Data File : VW030367.D
 Acq On : 30 Mar 2023 18:19
 Operator : SY/MD
 Sample : 02130-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C11B9

Quant Time: Mar 31 05:43:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 31 05:37:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	139490	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	135683	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	68389	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	34708	3.757	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.200%
7) Chloroethane-d5	1.536	69	38684	4.315	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.200%
11) 1,1-Dichloroethene-d2	2.063	63	63354	2.890	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.800%#
20) 2-Butanone-d5	3.828	46	135419	56.711	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.420%
24) Chloroform-d	4.259	84	105474	4.717	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	4.953	65	53904	5.110	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	4.969	84	185543	4.853	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	5.998	67	59587	4.978	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.252	98	155103	4.417	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
43) trans-1,3-Dichloroprop...	7.561	79	20045	4.448	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.037	63	97929	54.074	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.140%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	45137	4.784	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	11.570	152	64504	5.277	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.600%
Target Compounds						
13) Acetone	2.156	43	6765	2.963	ug/L	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
Data File : VV030367.D
Acq On : 30 Mar 2023 18:19
Operator : SY/MD
Sample : 02130-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C11B9

Quant Time: Mar 31 05:43:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 31 05:37:21 2023
Response via : Initial Calibration

