

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
 Data File : VV030368.D
 Acq On : 30 Mar 2023 18:43
 Operator : SY/MD
 Sample : 02130-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 31 05:43:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 31 05:37:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	137347	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	138058	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	65329	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	35891	3.946	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	79.000%	
7) Chloroethane-d5	1.536	69	39040	4.422	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.400%	
11) 1,1-Dichloroethene-d2	2.063	63	62559	2.898	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	58.000%#	
20) 2-Butanone-d5	3.815	46	123388	52.479	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.960%	
24) Chloroform-d	4.259	84	100942	4.585	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
26) 1,2-Dichloroethane-d4	4.953	65	52276	5.033	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.600%	
32) Benzene-d6	4.970	84	179813	4.622	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.400%	
36) 1,2-Dichloropropane-d6	5.998	67	56761	4.660	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.200%	
41) Toluene-d8	7.252	98	150210	4.204	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.000%	
43) trans-1,3-Dichloroprop...	7.561	79	18777	4.095	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.800%	
46) 2-Hexanone-d5	8.034	63	89839	48.754	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.500%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	42742	4.453	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.000%	
66) 1,2-Dichlorobenzene-d4	11.571	152	60035	5.142	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.800%	
Target Compounds						Qvalue
16) Methylene chloride	2.455	84	3399	0.250	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033023\
Data File : VV030368.D
Acq On : 30 Mar 2023 18:43
Operator : SY/MD
Sample : 02130-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 31 05:43:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 31 05:37:21 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.250 ug/L
RT: 2.455 min Scan# 44
Delta R.T. 0.003 min
Lab File: VV030368.D
Acq: 30 Mar 2023 18:43

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Tgt Ion: 84 Resp: 3399
Ion Ratio Lower Upper
84 100
86 63.1 44.0 81.6
49 132.9 90.6 168.3

