

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV033120\
 Data File : VV015304.D
 Acq On : 31 Mar 2020 11:19
 Operator : SY/MD
 Sample : VV0331WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK99

Quant Time: Apr 01 01:29:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 01 01:17:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	576053	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	558375	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	241166	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	176162	41.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.88%
7) Chloroethane-d5	1.58	69	178879	45.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.52%
11) 1,1-Dichloroethene-d2	2.12	63	286636	31.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.28%
21) 2-Butanone-d5	3.94	46	292257	106.40	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.40%
24) Chloroform-d	4.38	84	349930	46.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.68%
26) 1,2-Dichloroethane-d4	5.06	65	241928	49.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.04%
32) Benzene-d6	5.08	84	682240	46.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.62%
36) 1,2-Dichloropropane-d6	6.10	67	224662	49.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.58%
41) Toluene-d8	7.34	98	645192	45.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.50%
43) trans-1,3-Dichloropropene-	7.64	79	115816	46.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.32%
47) 2-Hexanone-d5	8.12	63	220462	103.10	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.10%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	320791	47.69	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.38%
64) 1,2-Dichlorobenzene-d4	11.65	152	224983	46.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV033120\
Data File : VV015304.D
Acq On : 31 Mar 2020 11:19
Operator : SY/MD
Sample : VV0331WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK99

Quant Time: Apr 01 01:29:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM033020WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 01 01:17:34 2020
Response via : Initial Calibration

