

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033121\
 Data File : VV020819.D
 Acq On : 31 Mar 2021 10:36
 Operator : SY/MD
 Sample : VSTD0.0536
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05302

Quant Time: Apr 01 04:04:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM033121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 01 04:03:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	5270	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	5181	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2832	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	222	0.04	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.066	65	195	0.05	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.069	67	192	0.04	ug/L	0.00
8) Toluene-d8	7.332	98	587	0.05	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.216	84	142	0.05	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	190	0.04	ug/L #	55
6) Trichloroethene	5.925	95	198	0.05	ug/L	91
9) 1,2-Dibromoethane	8.369	107	107	0.04	ug/L #	81
12) 1,2,3-Trichloropropane	10.267	75	117	0.05	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.427	75	28	0.07	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033121\
Data File : VV020819.D
Acq On : 31 Mar 2021 10:36
Operator : SY/MD
Sample : VSTD0.0536
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05302

Quant Time: Apr 01 04:04:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM033121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 01 04:03:39 2021
Response via : Initial Calibration

