

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033121\
 Data File : VV020823.D
 Acq On : 31 Mar 2021 12:39
 Operator : SY/MD
 Sample : VSTD00234
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002306

Quant Time: Apr 01 04:04:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM033121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 01 04:03:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	5091	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	5079	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2639	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	6945	1.18	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	7591	1.95	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	8328	1.82	ug/L	0.00
8) Toluene-d8	7.317	98	20860	1.65	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	5690	2.01	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	5001	0.98	ug/L	98
6) Trichloroethene	5.925	95	7150	1.78	ug/L	97
9) 1,2-Dibromoethane	8.353	107	4776	2.01	ug/L #	92
12) 1,2,3-Trichloropropane	10.267	75	4537	2.04	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.426	75	760	1.96	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033121\
Data File : VV020823.D
Acq On : 31 Mar 2021 12:39
Operator : SY/MD
Sample : VSTD00234
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002306

Quant Time: Apr 01 04:04:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM033121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 01 04:03:39 2021
Response via : Initial Calibration

