

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033121\
 Data File : VV020821.D
 Acq On : 31 Mar 2021 11:52
 Operator : SY/MD
 Sample : VSTD0.532
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5304

Quant Time: Apr 01 04:04:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM033121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 01 04:03:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	5360	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	5210	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2858	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	1879	0.30	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	2044	0.50	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	2183	0.46	ug/L	0.00
8) Toluene-d8	7.317	98	5617	0.43	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	1507	0.52	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	1361	0.25	ug/L	100
6) Trichloroethene	5.925	95	1835	0.45	ug/L	100
9) 1,2-Dibromoethane	8.368	107	1199	0.49	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	1140	0.47	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.426	75	193	0.46	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033121\
Data File : VV020821.D
Acq On : 31 Mar 2021 11:52
Operator : SY/MD
Sample : VSTD0.532
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5304

Quant Time: Apr 01 04:04:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM033121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 01 04:03:39 2021
Response via : Initial Calibration

