



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033123\
Data File : VV030372.D
Acq On : 31 Mar 2023 08:54
Operator : SY/MD
Sample : BFB241
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB241

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Title : TRACE VOA SFAM1.0
Last Update : Fri Mar 31 05:37:21 2023



AutoFind: Scans 2791, 2792, 2793; Background Corrected with Scan 2778

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	17315	PASS
75	95	30	80	52.5	45728	PASS
95	95	100	100	100.0	87064	PASS
96	95	5	9	6.4	5599	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	91.2	79381	PASS
175	174	5	9	8.4	6660	PASS
176	174	95	105	98.0	77765	PASS
177	176	5	10	6.8	5261	PASS

m/z	Abundance
35.60	116.0
36.60	125.0
38.10	109.0
39.00	158.0
40.10	233.0
41.00	147.0
42.00	136.0
42.90	136.0
44.10	256.0
48.00	108.0
52.50	102.0
57.10	119.0
64.60	106.0
71.90	115.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241

m/z	Abundance
36.00	792.0
37.10	4878.0
38.10	4080.0
39.10	2129.0
39.90	329.0
40.90	156.0
42.90	138.0
43.20	140.0
44.10	698.0
45.00	885.0
46.00	218.0
47.10	1098.0
48.00	662.0
49.00	3581.0
50.10	16744.0
51.10	5145.0
52.00	308.0
53.10	179.0
53.80	149.0
55.00	332.0
56.00	1546.0
57.10	2651.0
57.90	160.0
60.00	945.0
61.00	4377.0
62.00	4356.0
63.00	3935.0
64.00	597.0
65.10	856.0
65.80	154.0
67.10	253.0
68.00	8571.0
69.10	8503.0
70.00	620.0
72.00	593.0
73.00	3959.0
74.00	14580.0
75.10	44336.0
76.10	4011.0
77.00	517.0
78.00	443.0
78.90	3762.0
80.00	810.0
80.90	3678.0
81.90	786.0
85.90	220.0
87.00	3025.0
88.00	3870.0
90.90	394.0
92.00	2768.0
93.00	3631.0
94.10	10680.0
95.00	83600.0
96.10	5608.0
97.10	207.0
103.90	450.0
104.80	179.0
105.90	445.0
106.90	117.0
108.30	109.0
110.00	125.0
111.10	121.0
114.80	162.0
115.90	400.0
116.90	930.0
117.90	294.0
118.90	655.0
127.80	373.0
128.90	202.0
129.90	376.0
130.90	180.0
134.80	200.0
135.10	193.0
136.90	169.0
140.00	105.0
141.00	996.0
142.00	141.0
142.90	1488.0
144.90	554.0
145.90	178.0
147.90	232.0
149.80	127.0
154.80	243.0
157.10	169.0
161.10	115.0
170.10	133.0
170.70	219.0
171.00	203.0
172.00	808.0
174.00	74256.0
175.00	6320.0
176.00	73000.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241

177.004962.0

178.00171.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241

m/z	Abundance
36.00	921.0
37.10	5188.0
38.10	4393.0
39.10	2287.0
39.90	302.0
41.00	158.0
41.80	112.0
42.40	123.0
42.90	149.0
43.40	142.0
44.10	716.0
45.00	846.0
46.00	225.0
47.10	1161.0
48.00	722.0
49.10	3723.0
50.10	17864.0
51.10	5454.0
52.00	299.0
52.80	171.0
53.10	189.0
53.80	188.0
55.00	381.0
56.00	1599.0
57.10	2714.0
57.90	144.0
60.00	958.0
61.10	4839.0
62.00	4444.0
63.00	4287.0
64.00	639.0
65.00	874.0
65.90	159.0
67.10	289.0
68.00	9256.0
69.00	9087.0
70.00	652.0
72.00	644.0
73.00	4348.0
74.10	15821.0
75.10	47416.0
76.10	4314.0
77.00	544.0
78.00	517.0
79.00	3870.0
80.00	982.0
80.90	3880.0
81.90	824.0
83.20	102.0
85.00	111.0
85.90	188.0
87.00	3356.0
88.00	3928.0
90.90	435.0
92.00	2892.0
93.00	3794.0
94.10	11350.0
95.10	90248.0
96.10	5742.0
97.20	211.0
103.90	515.0
104.80	190.0
105.90	512.0
107.00	134.0
110.00	115.0
111.10	111.0
111.70	104.0
114.80	184.0
115.90	434.0
116.90	939.0
117.90	284.0
118.90	697.0
125.90	114.0
127.80	400.0
128.80	182.0
129.90	385.0
130.90	176.0
133.70	107.0
134.70	215.0
135.10	194.0
136.90	192.0
140.90	1165.0
141.90	159.0
143.00	1570.0
144.90	587.0
145.90	200.0
146.80	101.0
147.90	203.0
149.80	110.0
154.80	293.0
157.00	225.0
158.70	131.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241

160.90	161.0
170.20	132.0
170.70	263.0
171.10	237.0
172.00	873.0
174.00	81936.0
175.00	6870.0
176.00	80736.0
177.00	5481.0
178.00	172.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241

m/z	Abundance
36.00	1009.0
37.10	4729.0
38.10	4178.0
39.10	2093.0
39.90	286.0
41.00	156.0
42.20	122.0
42.90	141.0
44.00	759.0
45.00	732.0
45.90	231.0
47.10	1178.0
48.00	678.0
49.10	3661.0
50.10	17336.0
51.10	5306.0
52.00	251.0
53.10	210.0
53.80	186.0
55.00	383.0
56.00	1531.0
57.00	2399.0
57.90	135.0
60.00	940.0
61.10	4797.0
62.00	4380.0
63.00	4141.0
64.10	601.0
65.10	862.0
65.90	174.0
67.10	295.0
68.00	9226.0
69.00	8929.0
70.00	649.0
70.90	104.0
72.00	579.0
73.00	4234.0
74.10	15316.0
75.10	45432.0
76.10	4127.0
77.00	564.0
78.00	512.0
79.00	3696.0
79.90	1033.0
80.90	3895.0
81.90	819.0
83.20	102.0
85.00	111.0
86.00	182.0
87.00	3710.0
88.00	3558.0
91.00	443.0
92.00	2764.0
93.00	3743.0
94.10	10981.0
95.10	87344.0
96.10	5448.0
97.20	198.0
103.90	567.0
104.80	165.0
105.90	566.0
107.00	132.0
109.70	125.0
111.00	102.0
111.80	126.0
114.80	211.0
115.90	398.0
116.90	809.0
117.90	280.0
118.90	641.0
122.00	103.0
125.90	116.0
127.80	406.0
128.80	172.0
129.90	350.0
131.00	186.0
133.70	111.0
134.80	210.0
136.90	254.0
140.90	1214.0
141.80	167.0
143.00	1489.0
143.90	110.0
144.90	547.0
145.90	225.0
148.00	194.0
149.80	102.0
153.90	105.0
154.90	324.0
156.90	256.0
158.70	172.0
160.90	197.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241

170.20	117.0
170.70	300.0
172.00	918.0
174.00	81952.0
175.00	6790.0
176.00	79560.0
177.00	5341.0
178.00	169.0

Instrument :
MSVOA_V
ClientSampleId :
BFB241