

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040120\
 Data File : VV015336.D
 Acq On : 01 Apr 2020 17:49
 Operator : SY/MD
 Sample : L1968-17
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Apr 02 04:08:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 02 03:22:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	730289	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	696031	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	318262	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	273077	41.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.40%
7) Chloroethane-d5	1.58	69	237006	42.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.74%
11) 1,1-Dichloroethene-d2	2.12	63	387797	33.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.26%
21) 2-Butanone-d5	3.94	46	323660	93.00	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.00%
24) Chloroform-d	4.38	84	456425	42.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.44%
26) 1,2-Dichloroethane-d4	5.06	65	289917	43.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.10%
32) Benzene-d6	5.08	84	945572	44.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.22%
36) 1,2-Dichloropropane-d6	6.09	67	281805	44.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.18%
41) Toluene-d8	7.34	98	903105	44.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.12%
43) trans-1,3-Dichloropropene-	7.64	79	143341	41.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.10%
47) 2-Hexanone-d5	8.12	63	246869	85.53	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.53%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	396284	44.07	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.14%
64) 1,2-Dichlorobenzene-d4	11.65	152	332394	45.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040120\
Data File : VV015336.D
Acq On : 01 Apr 2020 17:49
Operator : SY/MD
Sample : L1968-17
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Apr 02 04:08:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 02 03:22:54 2020
Response via : Initial Calibration

