

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040120\
 Data File : VV015359.D
 Acq On : 02 Apr 2020 05:25
 Operator : SY/MD
 Sample : L2064-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0Y03

Quant Time: Apr 02 07:18:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 02 06:34:01 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	683714	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	641435	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	287932	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	228575	37.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.56%
7) Chloroethane-d5	1.58	69	206853	39.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.00%
11) 1,1-Dichloroethene-d2	2.12	63	343892	31.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.76%
21) 2-Butanone-d5	3.94	46	298148	91.50	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.50%
24) Chloroform-d	4.38	84	431191	42.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.20%
26) 1,2-Dichloroethane-d4	5.06	65	277796	44.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.12%
32) Benzene-d6	5.08	84	877736	44.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.86%
36) 1,2-Dichloropropane-d6	6.09	67	266153	45.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.40%
41) Toluene-d8	7.34	98	834569	44.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.36%
43) trans-1,3-Dichloropropene-	7.64	79	133611	41.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.04%
47) 2-Hexanone-d5	8.12	63	232838	87.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.54%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	376467	45.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.86%
64) 1,2-Dichlorobenzene-d4	11.65	152	308818	47.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.30%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.76	101	113607	9.845 ug/L	99
25) Chloroform	4.40	83	12215	1.325 ug/L	79
34) Trichloroethene	5.94	95	6641	1.321 ug/L	96
46) Tetrachloroethene	8.00	164	191863	53.237 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040120\
Data File : VV015359.D
Acq On : 02 Apr 2020 05:25
Operator : SY/MD
Sample : L2064-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0Y03

Quant Time: Apr 02 07:18:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 02 06:34:01 2020
Response via : Initial Calibration

