

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040120\
 Data File : VV015338.D
 Acq On : 01 Apr 2020 18:37
 Operator : SY/MD
 Sample : L1976-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Apr 02 04:54:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 02 03:22:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	708186	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	667579	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	308131	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	256377	40.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.74%
7) Chloroethane-d5	1.58	69	225900	41.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.30%
11) 1,1-Dichloroethene-d2	2.12	63	363907	32.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.12%
21) 2-Butanone-d5	3.93	46	301132	89.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.22%
24) Chloroform-d	4.38	84	435359	41.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.06%
26) 1,2-Dichloroethane-d4	5.06	65	271167	41.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.06%
32) Benzene-d6	5.07	84	888796	43.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.46%
36) 1,2-Dichloropropane-d6	6.09	67	268063	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.44%
41) Toluene-d8	7.34	98	842334	42.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.70%
43) trans-1,3-Dichloropropene-	7.64	79	132544	39.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.14%
47) 2-Hexanone-d5	8.11	63	238003	85.97	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.97%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	378473	43.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.76%
64) 1,2-Dichlorobenzene-d4	11.65	152	321250	45.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.66%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040120\
Data File : VV015338.D
Acq On : 01 Apr 2020 18:37
Operator : SY/MD
Sample : L1976-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Apr 02 04:54:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 02 03:22:54 2020
Response via : Initial Calibration

