

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040121\
 Data File : VV020856.D
 Acq On : 02 Apr 2021 02:21
 Operator : SY/MD
 Sample : M1845-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :


Quant Time: Apr 02 08:33:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 02 08:30:00 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	597848	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	575076	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	274964	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	141849	33.42	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	66.84%		
7) Chloroethane-d5	1.571	69	132666	41.34	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	82.68%		
11) 1,1-Dichloroethene-d2	2.108	63	250519	31.28	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.56%		
21) 2-Butanone-d5	3.899	46	232358	84.15	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	84.15%		
24) Chloroform-d	4.352	84	412967	48.95	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.90%		
26) 1,2-Dichloroethane-d4	5.037	65	275882	49.64	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.28%		
32) Benzene-d6	5.053	84	742113	48.41	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.82%		
36) 1,2-Dichloropropane-d6	6.072	67	222204	47.22	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	94.44%		
41) Toluene-d8	7.320	98	683841	47.93	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.86%		
43) trans-1,3-Dichloroprop...	7.625	79	114599	45.77	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	91.54%		
47) 2-Hexanone-d5	8.092	63	173350	87.86	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	87.86%		
57) 1,1,2,2-Tetrachloroeth...	10.220	84	302557	44.14	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	88.28%		
64) 1,2-Dichlorobenzene-d4	11.632	152	284818	53.37	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	106.74%		
Target Compounds						Qvalue
5) Vinyl chloride	1.314	62	10852	2.38	ug/L #	69
10) 1,1,2-Trichloro-1,2,2-...	2.111	101	2633	0.70	ug/L #	18
17) trans-1,2-Dichloroethene	2.767	96	5985	1.50	ug/L	91
20) cis-1,2-Dichloroethene	3.918	96	39668	9.27	ug/L	93
34) Trichloroethene	5.921	95	109443	25.17	ug/L	95
48) 2-Hexanone	8.092	43	20423	4.80	ug/L #	71
59) 1,2,3-Trichloropropane	11.252	75	24807	4.57	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 02 08:33:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 02 08:30:00 2021
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 2.38 ug/L
 RT: 1.314 min Scan# 85
 Delta R.T. 0.003 min
 Lab File: VV020856.D
 Acq: 02 Apr 2021 02:21




Instrument :
 MSVOA_V
 ClientSampled :


Tgt Ion: 62 Resp: 10852
 Ion Ratio Lower Upper
 62 100
 64 51.1 23.3 43.3#



#10
 1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 0.70 ug/L
 RT: 2.111 min Scan# 333
 Delta R.T. -0.006 min
 Lab File: VV020856.D
 Acq: 02 Apr 2021 02:21


Tgt Ion: 101 Resp: 2633
 Ion Ratio Lower Upper
 101 100
 85 1.6 35.4 53.0#
 151 0.0 63.5 95.3#








#17
trans-1,2-Dichloroethene
Concen: 1.50 ug/L
RT: 2.767 min Scan# 537
Delta R.T. 0.003 min
Lab File: VV020856.D
Acq: 02 Apr 2021 02:21

Instrument :
MSVOA_V
ClientSampled :







Tgt Ion: 96 Resp: 5985
Ion Ratio Lower Upper
96 100
61 126.3 97.9 181.9
98 65.8 43.9 81.5



#20
cis-1,2-Dichloroethene
Concen: 9.27 ug/L
RT: 3.918 min Scan# 895
Delta R.T. 0.000 min
Lab File: VV020856.D
Acq: 02 Apr 2021 02:21


Tgt Ion: 96 Resp: 39668
Ion Ratio Lower Upper
96 100
61 120.0 89.9 167.0
68 0.0 0.0 0.0








#34
Trichloroethene
Concen: 25.17 ug/L
RT: 5.921 min Scan# 1518
Delta R.T. 0.003 min
Lab File: VV020856.D
Acq: 02 Apr 2021 02:21

Instrument :
MSVOA_V
ClientSampled :




Tgt Ion: 95 Resp: 109443

Ion	Ratio	Lower	Upper
95	100		
97	64.1	43.5	80.9
132	101.3	69.5	129.1
130	109.4	70.7	131.3






#48
2-Hexanone
Concen: 4.80 ug/L
RT: 8.092 min Scan# 2193
Delta R.T. -0.051 min
Lab File: VV020856.D
Acq: 02 Apr 2021 02:21



Tgt Ion: 43 Resp: 20423

Ion	Ratio	Lower	Upper
43	100		
58	34.5	45.8	68.8#
57	29.4	15.4	23.2#
100	0.4	10.2	15.4#





#59
 1,2,3-Trichloropropane
 Concen: 4.57 ug/L
 RT: 11.252 min Scan# 3176
 Delta R.T. 0.974 min
 Lab File: VV020856.D
 Acq: 02 Apr 2021 02:21

Instrument :
 MSVOA_V
 ClientSampleId :

Tgt Ion: 75 Resp: 24807
 Ion Ratio Lower Upper
 75 100
 77 35.7 25.3 37.9

