

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040121\
 Data File : VV020838.D
 Acq On : 01 Apr 2021 14:39
 Operator : SY/MD
 Sample : M1845-01RE
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1B1RE

Quant Time: Apr 02 07:18:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 02 07:14:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	488748	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	476429	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	231487	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	173172	49.91	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 99.82%			
7) Chloroethane-d5	1.571	69	149281	56.91	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 113.82%			
11) 1,1-Dichloroethene-d2	2.108	63	290395	44.36	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 88.72%			
21) 2-Butanone-d5	3.896	46	256163	113.47	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 113.47%			
24) Chloroform-d	4.352	84	441515	64.02	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 128.04%#			
26) 1,2-Dichloroethane-d4	5.037	65	299804	65.99	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 131.98%#			
32) Benzene-d6	5.053	84	806171	63.48	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 126.96%#			
36) 1,2-Dichloropropane-d6	6.072	67	235975	60.53	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 121.06%#			
41) Toluene-d8	7.320	98	738675	62.49	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 124.98%#			
43) trans-1,3-Dichloroprop...	7.625	79	124037	59.79	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 119.58%			
47) 2-Hexanone-d5	8.092	63	184377	112.80	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 112.80%			
57) 1,1,2,2-Tetrachloroeth...	10.220	84	329425	58.01	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 116.02%			
64) 1,2-Dichlorobenzene-d4	11.632	152	301536	67.12	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 134.24%#			
Target Compounds						
20) cis-1,2-Dichloroethene	3.921	96	3432	0.98	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040121\
Data File : VV020838.D
Acq On : 01 Apr 2021 14:39
Operator : SY/MD
Sample : M1845-01RE
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1B1RE

Quant Time: Apr 02 07:18:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 02 07:14:26 2021
Response via : Initial Calibration

