

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040121\
 Data File : VW020843.D
 Acq On : 01 Apr 2021 16:38
 Operator : SY/MD
 Sample : M1879-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1A1

Quant Time: Apr 02 07:19:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 02 07:14:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	692887	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	662903	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	323604	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	169806	34.52	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 69.04%			
7) Chloroethane-d5	1.571	69	148995	40.06	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 80.12%			
11) 1,1-Dichloroethene-d2	2.108	63	288818	31.12	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 62.24%			
21) 2-Butanone-d5	3.896	46	269692	84.27	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 84.27%			
24) Chloroform-d	4.352	84	442143	45.22	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 90.44%			
26) 1,2-Dichloroethane-d4	5.037	65	302854	47.02	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 94.04%			
32) Benzene-d6	5.053	84	807965	45.73	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 91.46%			
36) 1,2-Dichloropropane-d6	6.072	67	238465	43.96	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 87.92%			
41) Toluene-d8	7.320	98	735369	44.71	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 89.42%			
43) trans-1,3-Dichloroprop...	7.625	79	126842	43.95	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 87.90%			
47) 2-Hexanone-d5	8.092	63	194227	85.40	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 85.40%			
57) 1,1,2,2-Tetrachloroeth...	10.220	84	332940	42.13	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 84.26%			
64) 1,2-Dichlorobenzene-d4	11.632	152	306611	48.82	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 97.64%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040121\
Data File : VV020843.D
Acq On : 01 Apr 2021 16:38
Operator : SY/MD
Sample : M1879-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1A1

Quant Time: Apr 02 07:19:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 02 07:14:26 2021
Response via : Initial Calibration

